

Certificate of analysis

Product Name/Description: Western BLoT Quant HRP Substrate
Cat. #: T7102A
Lot #: Specified on product label
Storage Conditions: 4 degrees C
Expiration Date: 1 year from date of receipt under proper storage condition

Package Size: 100 ml

Package Contents:

1. Western BLoT Quant Luminol/Enhancer Solution	50 ml
2. Western BLoT Quant Peroxide Reagent	50 ml

Quality Control Data:

The reagents are tested by incubation for 2 minutes with nitrocellulose strips containing eight serial dilutions of secondary antibody conjugated with HRP in the range from 5 ng to 40 pg followed by imaging in a CCD digital imager.

In standard imaging conditions, at least seven bands should be detected.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

Notice To Purchaser:

Please refer to the website.

Product Documents, Patent Markings, Trademarks, & Safety Information:

Refer to website for availability to download or contact your local office:

www.takara-bio.co.jp (Japanese)

www.takarabio.com (English)

Licence Statements:

Please visit our website and click the "license" button on the product's web page.

©2019 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.